



QUASIOPTICAL SOURCE

QS-1100 (OV-83) SN 127940

I. TECHNICAL DATA

1. OPERATING RANGE, GHz 910...1100
2. CATHODE VOLTAGE (NEGATIVE), V 3500...5500
3. CATHODE CURRENT, mA 22
4. MAGNETIC FIELD STRENGTH (min), Oe..... 13000
5. HEATER VOLTAGE (alternative), V 6.3
6. HEATER CURRENT, A 1.65
7. OUTPUT POWER, mW up to 5


II. CONTROL POINT PARAMETERS

Cathode current is varies with Cathode voltage and Heater current settings. The control point parameters are defined to prevent cathode damage due to excessive current

1. Cathode voltage, V	4000
2. Cathode current, mA	22
3. Heater current, A	1.65
5. Output frequency, GHz	973


If cathode current at the control cathode voltage exceeds specified value, reduce the heater current to prevent cathode degradation. BWO can not be repaired if the cathode is damaged.

III. CALIBRATION CURVE



IV. OUTPUT POWER PATTERN

Output spectrum varies with alignment of the tube in a magnet and heater current. Typical data for an optimal alignment in MS-1.2 magnetic system are shown below. Please note that lifetime of the tube is an exponential function of the heater current. Please use lower heater current for preliminary measurements and increase the heater current on as needed basis to extend operating lifetime of the tube.



Turn ON Procedure for QS1-1100 (Using VR6-M BWO Power supply)

1. Confirm the following settings for switches on VR6-M front panel
 - “Range kV” selector is set at **6 kV**
 - “Control” switch is in “INT” position
 - “Heater current control” is set at 2-3% below the value specified in control point parameter table (Section II)
2. Turn ON the main power by pushing the **big black “ON” button on the right side of VR6-M**. Green Power indicator will light up at this point.
3. Press the “LOAD” button (above the heater current control). Green indicator above the load button will start blinking and heater current will increase gradually. Once the heater current reaches a preset value, the green indicator will stop blinking.
4. Turn the “COARSE” adjustment knob all the way counter clockwise (setting it to zero) and make 2 turns in the clockwise direction. This will preset the high voltage in 1000-1500 V range
5. Switch ON high voltage by pressing the **big black “ON” button on the left side of VR6-M** front panel. Red “High Voltage” indicator will light up and cathode voltage meter should show 500-1000V at this point.
6. Set cathode voltage at the control point value (specified in Section II) by “COARSE” potentiometer.
7. By adjusting the heater current set the cathode current to a value 1-2 mA smaller than the control point value. The heater current setting is adjusted by entering a new value and pressing the load button again. Once the tube warms up (after 20-30 minutes), the cathode current should reach the control point value.
8. Change the cathode voltage to choose the output frequency required. Do not exceed maximum voltage specified in Section I.

Turn OFF procedure for QS1-1100 (Using VR6-M BWO Power supply)

1. **Turn OFF the high voltage** by pressing the big red button on the left side of VR6-M front panel.
2. Press the red “UNLOAD” button. Wait until the heater current goes down to zero and green indicator stops blinking
4. Switch OFF the main power by pressing the big red button on the right side of VR6-M front panel.

Please follow recommended Turn ON/OFF procedures to prevent damage of quasi-optical sources.

Please check the cathode current at the control point once every 30 minutes to make sure it does not exceed the value specified in Section II.